

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038077.D
 Acq On : 04 Oct 2023 19:52
 Operator : JC/MD
 Sample : 04622-02ME
 Misc : 3.36g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-4(0-5)ME

Quant Time: Oct 05 01:07:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175512	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73516	47.533	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.060%
35) Dibromofluoromethane	5.391	113	51445	42.725	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.460%
50) Toluene-d8	8.647	98	220883	47.934	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.860%
62) 4-Bromofluorobenzene	11.085	95	87750	52.324	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.640%
Target Compounds						
					Qvalue	
40) Benzene	6.031	78	29676	5.928	ug/l	95
52) Toluene	8.720	92	66913	20.653	ug/l	92
67) Ethyl Benzene	10.195	91	83237	13.026	ug/l	100
68) m/p-Xylenes	10.299	106	65061	26.587	ug/l	100
69) o-Xylene	10.640	106	34451	14.825	ug/l	100
70) Styrene	10.659	104	35085	9.141	ug/l	91
73) Isopropylbenzene	10.963	105	9949	1.469	ug/l	97
78) n-propylbenzene	11.305	91	16917	2.033	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	63170	10.961	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	162236	27.912	ug/l	98
86) p-Isopropyltoluene	12.012	119	15345	2.522	ug/l	94
89) n-Butylbenzene	12.323	91	18980	3.125	ug/l #	26

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038077.D
Acq On : 04 Oct 2023 19:52
Operator : JC/MD
Sample : 04622-02ME
Misc : 3.36g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-4(0-5)ME

Quant Time: Oct 05 01:07:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

