


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038051.D
Acq On : 04 Oct 2023 08:33
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Title : SW846 8260
Last Update : Wed Oct 04 09:08:35 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	8253	PASS
75	95	30	60	55.2	21197	PASS
95	95	100	100	100.0	38376	PASS
96	95	5	9	7.1	2717	PASS
173	174	0.00	2	1.2	358	PASS
174	95	50	100	77.8	29864	PASS
175	174	5	9	7.5	2248	PASS
176	174	95	101	95.5	28515	PASS
177	176	5	9	7.1	2037	PASS