

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043273.D
 Acq On : 04 Oct 2024 09:01
 Operator : JC/MD
 Sample : VX1004MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1004MBL01

Quant Time: Oct 04 23:23:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	149302	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58428	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68523	51.836	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.680%
35) Dibromofluoromethane	5.379	113	50297	48.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	8.647	98	178891	50.135	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.079	95	65005	50.146	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043273.D
Acq On : 04 Oct 2024 09:01
Operator : JC/MD
Sample : VX1004MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1004MBL01

Quant Time: Oct 04 23:23:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

