

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043274.D
 Acq On : 04 Oct 2024 09:24
 Operator : JC/MD
 Sample : VX1004WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1004WBL01

Quant Time: Oct 04 23:24:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	77627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144088	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53080	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66194	52.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.420%
35) Dibromofluoromethane	5.379	113	47988	48.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.647	98	170850	49.614	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.220%
62) 4-Bromofluorobenzene	11.079	95	59917	47.894	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043274.D
Acq On : 04 Oct 2024 09:24
Operator : JC/MD
Sample : VX1004WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1004WBL01

Quant Time: Oct 04 23:24:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

