

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043277.D
 Acq On : 04 Oct 2024 10:37
 Operator : JC/MD
 Sample : P4263-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20240930DL

Quant Time: Oct 04 23:26:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	82634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	157373	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70420	52.181	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.360%			
35) Dibromofluoromethane	5.385	113	52018	47.929	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.860%			
50) Toluene-d8	8.647	98	184787	49.131	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.260%			
62) 4-Bromofluorobenzene	11.079	95	65619	48.024	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.040%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	1760	1.904	ug/l	96
44) Trichloroethene	7.123	130	39817	36.112	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043277.D
Acq On : 04 Oct 2024 10:37
Operator : JC/MD
Sample : P4263-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20240930DL

Quant Time: Oct 04 23:26:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

