

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043278.D
 Acq On : 04 Oct 2024 11:00
 Operator : JC/MD
 Sample : P4263-04DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D2-20240930DL

Quant Time: Oct 04 23:26:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	148561	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66997	53.117	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.240%	
35) Dibromofluoromethane	5.385	113	49024	47.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.700%	
50) Toluene-d8	8.647	98	178695	50.330	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.660%	
62) 4-Bromofluorobenzene	11.079	95	63350	49.114	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.220%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	1379	1.596	ug/l	92
12) 1,1-Dichloroethene	2.313	96	828	0.965	ug/l	97
44) Trichloroethene	7.123	130	32974	31.680	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043278.D
Acq On : 04 Oct 2024 11:00
Operator : JC/MD
Sample : P4263-04DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D2-20240930DL

Quant Time: Oct 04 23:26:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

