

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043281.D
 Acq On : 04 Oct 2024 12:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 05 04:11:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	156222	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71141	53.369	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.740%
35) Dibromofluoromethane	5.385	113	51102	47.432	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.860%
50) Toluene-d8	8.647	98	182545	48.893	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.780%
62) 4-Bromofluorobenzene	11.079	95	65519	48.304	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.600%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1039	1.866	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043281.D
Acq On : 04 Oct 2024 12:09
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 05 04:11:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

