

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043285.D
 Acq On : 04 Oct 2024 13:42
 Operator : JC/MD
 Sample : P4263-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT162S1-20241001

Quant Time: Oct 04 23:30:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80551	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	152377	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69394	52.750	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.500%
35) Dibromofluoromethane	5.385	113	50014	47.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.647	98	181251	49.771	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.079	95	62302	47.091	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.180%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	406	0.454	ug/l #	60
16) Acetone	2.392	43	976	1.777	ug/l #	71
44) Trichloroethene	7.123	130	12223	11.449	ug/l	94
64) Tetrachloroethene	9.281	164	857	0.966	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043285.D
Acq On : 04 Oct 2024 13:42
Operator : JC/MD
Sample : P4263-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT162S1-20241001

Quant Time: Oct 04 23:30:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

