

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043286.D
 Acq On : 04 Oct 2024 14:05
 Operator : JC/MD
 Sample : P4263-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP11-20240930

Quant Time: Oct 04 23:30:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	81103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	157899	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71013	53.614	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.220%	
35) Dibromofluoromethane	5.385	113	51075	46.904	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.800%	
50) Toluene-d8	8.647	98	186697	49.474	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.940%	
62) 4-Bromofluorobenzene	11.079	95	65680	47.909	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.820%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1489	1.578	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.319	101	19268	21.234	ug/l	98
12) 1,1-Dichloroethene	2.312	96	2976	3.303	ug/l	98
16) Acetone	2.386	43	921	1.665	ug/l #	77
27) cis-1,2-Dichloroethene	4.489	96	6821	6.099	ug/l	87
30) Chloroform	5.086	83	4224	2.217	ug/l #	80
38) Carbon Tetrachloride	5.678	117	3050	1.873	ug/l	87
44) Trichloroethene	7.123	130	402279	363.635	ug/l	95
55) 1,1,2-Trichloroethane	9.159	97	1013	0.975	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043286.D
Acq On : 04 Oct 2024 14:05
Operator : JC/MD
Sample : P4263-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP11-20240930

Quant Time: Oct 04 23:30:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

