

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043293.D
 Acq On : 04 Oct 2024 16:47
 Operator : JC/MD
 Sample : P4263-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20240930

Quant Time: Oct 04 23:33:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	153282	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69479	53.002	ug/l	0.00
Spiked Amount 50.000	Range	74 - 125	Recovery	=	106.000%	
35) Dibromofluoromethane	5.385	113	50320	47.602	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	95.200%	
50) Toluene-d8	8.647	98	183254	50.024	ug/l	0.00
Spiked Amount 50.000	Range	86 - 113	Recovery	=	100.040%	
62) 4-Bromofluorobenzene	11.079	95	65340	49.096	ug/l	0.00
Spiked Amount 50.000	Range	77 - 121	Recovery	=	98.200%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	1504	1.611	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	22901	25.500	ug/l	98
12) 1,1-Dichloroethene	2.313	96	8461	9.488	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	3369	3.044	ug/l	91
30) Chloroform	5.086	83	3736	1.982	ug/l #	82
38) Carbon Tetrachloride	5.672	117	4421	2.796	ug/l	97
44) Trichloroethene	7.123	130	509100	474.055	ug/l	94
55) 1,1,2-Trichloroethane	9.159	97	1188	1.178	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043293.D
Acq On : 04 Oct 2024 16:47
Operator : JC/MD
Sample : P4263-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20240930

Quant Time: Oct 04 23:33:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

