

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043296.D
 Acq On : 04 Oct 2024 17:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 04 23:37:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	50.000	49.677	0.6	86	0.00
3 P	Chloromethane	50.000	47.068	5.9	84	0.00
4 C	Vinyl Chloride	50.000	49.012	2.0#	87	0.00
5 T	Bromomethane	50.000	49.860	0.3	91	0.00
6 T	Chloroethane	50.000	53.889	-7.8	95	0.00
7 T	Trichlorofluoromethane	50.000	49.670	0.7	91	0.00
8 T	Diethyl Ether	50.000	52.187	-4.4	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.948	-3.9	93	0.00
10 T	Methyl Iodide	50.000	50.034	-0.1	87	0.00
11 T	Tert butyl alcohol	250.000	294.277	-17.7	102	0.00
12 CM	1,1-Dichloroethene	50.000	48.650	2.7#	88	0.00
13 T	Acrolein	250.000	263.885	-5.6	94	0.00
14 T	Allyl chloride	50.000	50.699	-1.4	90	0.00
15 T	Acrylonitrile	250.000	269.257	-7.7	95	0.00
16 T	Acetone	250.000	237.483	5.0	91	0.00
17 T	Carbon Disulfide	50.000	43.929	12.1	80	0.00
18 T	Methyl Acetate	50.000	57.810	-15.6	99	0.00
19 T	Methyl tert-butyl Ether	50.000	53.261	-6.5	95	0.00
20 T	Methylene Chloride	50.000	49.020	2.0	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.980	0.0	89	0.00
22 T	Diisopropyl ether	50.000	53.487	-7.0	95	0.00
23 T	Vinyl Acetate	250.000	278.300	-11.3	94	0.00
24 P	1,1-Dichloroethane	50.000	53.213	-6.4	95	0.00
25 T	2-Butanone	250.000	267.226	-6.9	95	0.00
26 T	2,2-Dichloropropane	50.000	50.940	-1.9	92	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.556	-3.1	93	0.00
28 T	Bromochloromethane	50.000	49.461	1.1	92	0.00
29 T	Tetrahydrofuran	250.000	268.292	-7.3	98	0.00
30 C	Chloroform	50.000	52.124	-4.2#	95	0.00
31 T	Cyclohexane	50.000	49.010	2.0	89	0.00
32 T	1,1,1-Trichloroethane	50.000	51.759	-3.5	92	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.064	-2.1	91	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	50.246	-0.5	92	0.00
36 T	1,1-Dichloropropene	50.000	50.134	-0.3	91	0.00
37 T	Ethyl Acetate	50.000	52.038	-4.1	95	0.00
38 T	Carbon Tetrachloride	50.000	50.800	-1.6	90	0.00
39 T	Methylcyclohexane	50.000	49.591	0.8	89	0.00
40 TM	Benzene	50.000	50.880	-1.8	93	0.00
41 T	Methacrylonitrile	50.000	54.517	-9.0	93	0.00
42 TM	1,2-Dichloroethane	50.000	52.074	-4.1	93	0.00
43 T	Isopropyl Acetate	50.000	52.777	-5.6	93	0.00
44 TM	Trichloroethene	50.000	51.120	-2.2	95	0.00
45 C	1,2-Dichloropropane	50.000	51.701	-3.4#	93	0.00
46 T	Dibromomethane	50.000	52.352	-4.7	94	0.00
47 T	Bromodichloromethane	50.000	51.824	-3.6	91	0.00
48 T	Methyl methacrylate	50.000	52.978	-6.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
 Data File : VX043296.D
 Acq On : 04 Oct 2024 17:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 04 23:37:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1160.869	-16.1	102	0.00
50 S	Toluene-d8	50.000	49.725	0.5	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.770	-7.5	94	0.00
52 CM	Toluene	50.000	51.109	-2.2#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	52.474	-4.9	89	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.591	-3.2	91	0.00
55 T	1,1,2-Trichloroethane	50.000	52.104	-4.2	91	0.00
56 T	Ethyl methacrylate	50.000	54.647	-9.3	92	0.00
57 T	1,3-Dichloropropane	50.000	51.664	-3.3	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	275.486	-10.2	91	0.00
59 T	2-Hexanone	250.000	271.279	-8.5	94	0.00
60 T	Dibromochloromethane	50.000	51.982	-4.0	88	0.00
61 T	1,2-Dibromoethane	50.000	51.899	-3.8	91	0.00
62 S	4-Bromofluorobenzene	50.000	50.062	-0.1	87	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	87	0.00
64 T	Tetrachloroethene	50.000	51.276	-2.6	90	0.00
65 PM	Chlorobenzene	50.000	50.755	-1.5	89	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.183	-6.4	90	0.00
67 C	Ethyl Benzene	50.000	52.626	-5.3#	90	0.00
68 T	m/p-Xylenes	100.000	102.292	-2.3	88	0.00
69 T	o-Xylene	50.000	51.597	-3.2	90	0.00
70 T	Styrene	50.000	53.359	-6.7	90	0.00
71 P	Bromoform	50.000	53.164	-6.3	88	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	87	0.00
73 T	Isopropylbenzene	50.000	51.671	-3.3	89	0.00
74 T	N-amyl acetate	50.000	52.482	-5.0	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.873	-3.7	92	0.00
76 T	1,2,3-Trichloropropane	50.000	51.005	-2.0	92	0.00
77 T	Bromobenzene	50.000	50.618	-1.2	90	0.00
78 T	n-propylbenzene	50.000	52.666	-5.3	89	0.00
79 T	2-Chlorotoluene	50.000	50.653	-1.3	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.443	-2.9	90	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.731	-3.5	92	0.00
82 T	4-Chlorotoluene	50.000	51.161	-2.3	91	0.00
83 T	tert-Butylbenzene	50.000	51.190	-2.4	90	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.473	-2.9	90	0.00
85 T	sec-Butylbenzene	50.000	52.138	-4.3	90	0.00
86 T	p-Isopropyltoluene	50.000	52.614	-5.2	91	0.00
87 T	1,3-Dichlorobenzene	50.000	51.527	-3.1	92	0.00
88 T	1,4-Dichlorobenzene	50.000	49.042	1.9	90	0.00
89 T	n-Butylbenzene	50.000	53.915	-7.8	91	0.00
90 T	Hexachloroethane	50.000	51.218	-2.4	87	0.00
91 T	1,2-Dichlorobenzene	50.000	50.211	-0.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.277	-4.6	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.767	-3.5	88	0.00
94 T	Hexachlorobutadiene	50.000	52.268	-4.5	88	0.00
95 T	Naphthalene	50.000	51.940	-3.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100424\
Data File : VX043296.D
Acq On : 04 Oct 2024 17:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 04 23:37:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.626	-3.3	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6