

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100518\
 Data File : VX005071.D
 Acq On : 05 Oct 2018 17:31
 Operator : JC/MD
 Sample : VX1005MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1005MBL01

Quant Time: Oct 05 18:31:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	199438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	327186	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	312402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176345	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154231	54.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.28%	
35) Dibromofluoromethane	5.51	113	133615	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
50) Toluene-d8	8.72	98	453498	49.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.38%	
62) 4-Bromofluorobenzene	11.14	95	166295	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100518\
Data File : VX005071.D
Acq On : 05 Oct 2018 17:31
Operator : JC/MD
Sample : VX1005MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1005MBL01

Quant Time: Oct 05 18:31:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

