

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100518\
 Data File : VX005075.D
 Acq On : 05 Oct 2018 19:18
 Operator : JC/MD
 Sample : J5192-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-05-100218-A

Quant Time: Oct 05 19:46:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	363367	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	360604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	171315	53.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.02%	
35) Dibromofluoromethane	5.50	113	140249	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
50) Toluene-d8	8.72	98	523225	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.14	95	193752	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
Target Compounds						
16) Acetone	2.45	43	32021	16.970	ug/l	92
18) Methyl Acetate	2.78	43	6792	1.421	ug/l	99
20) Methylene Chloride	2.86	84	6158	1.429	ug/l	92
25) 2-Butanone	4.71	43	6589	2.369	ug/l	92
29) Tetrahydrofuran	5.17	42	2145	1.241	ug/l	96
30) Chloroform	5.21	83	3316	0.653	ug/l #	75
43) Isopropyl Acetate	6.47	43	198777	24.693	ug/l	99
95) Naphthalene	13.83	128	4145	0.960	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100518\
Data File : VX005075.D
Acq On : 05 Oct 2018 19:18
Operator : JC/MD
Sample : J5192-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
AU-05-100218-A

Quant Time: Oct 05 19:46:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

