

Data File : VX005079.D
Acq On : 05 Oct 2018 21:05
Operator : JC/MD
Sample : J5207-13DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW510-3236DL

Quant Time: Oct 06 00:14:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	224283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	352389	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	347389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196996	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	170759	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
35) Dibromofluoromethane	5.51	113	150383	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	8.72	98	509203	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	11.14	95	186973	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
Target Compounds						
6) Chloroethane	1.73	64	18743	7.925	ug/l	95
12) 1,1-Dichloroethene	2.37	96	7624	3.137	ug/l #	78
16) Acetone	2.45	43	2859	1.528	ug/l	96
24) 1,1-Dichloroethane	3.70	63	245483	46.090	ug/l	99
32) 1,1,1-Trichloroethane	5.50	97	5198	1.264	ug/l #	49
40) Benzene	6.15	78	6159	0.483	ug/l	91
85) sec-Butylbenzene	11.94	105	10209	0.820	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX005079.D
Acq On : 05 Oct 2018 21:05
Operator : JC/MD
Sample : J5207-13DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW510-3236DL

Quant Time: Oct 06 00:14:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

