

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100518\
 Data File : VX005093.D
 Acq On : 06 Oct 2018 04:12
 Operator : JC/MD
 Sample : VX1005WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1005WBL02

Quant Time: Oct 06 06:58:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	216228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	322909	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	317452	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	179937	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	163363	52.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.78%	
35) Dibromofluoromethane	5.50	113	146358	53.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.96%	
50) Toluene-d8	8.72	98	465539	51.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.34%	
62) 4-Bromofluorobenzene	11.14	95	167841	51.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100518\
Data File : VX005093.D
Acq On : 06 Oct 2018 04:12
Operator : JC/MD
Sample : VX1005WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1005WBL02

Quant Time: Oct 06 06:58:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

