

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100518\
 Data File : VX005098.D
 Acq On : 06 Oct 2018 06:26
 Operator : JC/MD
 Sample : J5207-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW511-1115

Quant Time: Oct 08 13:31:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	322668	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	180814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	168771	54.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.28%	
35) Dibromofluoromethane	5.50	113	147757	54.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.06%	
50) Toluene-d8	8.72	98	461874	50.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.14	95	167820	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100518\
Data File : VX005098.D
Acq On : 06 Oct 2018 06:26
Operator : JC/MD
Sample : J5207-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW511-1115

Quant Time: Oct 08 13:31:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

