

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100518\
 Data File : VX005101.D
 Acq On : 06 Oct 2018 07:46
 Operator : JC/MD
 Sample : J5224-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA18-QBC5

Quant Time: Oct 08 07:07:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	222515	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	325192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	172188	54.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.34%	
35) Dibromofluoromethane	5.50	113	143134	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
50) Toluene-d8	8.71	98	470942	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	11.14	95	172364	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
Target Compounds						
4) Vinyl Chloride	1.40	62	2452	0.710	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	8453	2.867	ug/l	92
37) Ethyl Acetate	4.87	43	3095	0.620	ug/l #	93
44) Trichloroethene	7.21	130	9295	3.212	ug/l	96
64) Tetrachloroethene	9.34	164	9593	3.013	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100518\
Data File : VX005101.D
Acq On : 06 Oct 2018 07:46
Operator : JC/MD
Sample : J5224-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA18-QBC5

Quant Time: Oct 08 07:07:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

