

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024567.D
 Acq On : 05 Oct 2021 13:42
 Operator : JC/MD
 Sample : M4063-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-03-(SA)

Quant Time: Oct 05 16:42:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	160313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240638	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113270	50.13	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.26%			
35) Dibromofluoromethane	5.397	113	93195	58.30	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 116.60%			
50) Toluene-d8	8.653	98	311962	53.02	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 106.04%			
62) 4-Bromofluorobenzene	11.085	95	118063	51.03	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.06%			
Target Compounds					Qvalue	
40) Benzene	6.049	78	2719	0.41	ug/l #	77
52) Toluene	8.726	92	1216	0.29	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024567.D
Acq On : 05 Oct 2021 13:42
Operator : JC/MD
Sample : M4063-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-03-(SA)

Quant Time: Oct 05 16:42:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

