

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024575.D
 Acq On : 05 Oct 2021 16:49
 Operator : JC/MD
 Sample : M4080-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P287-DITCH

Quant Time: Oct 06 06:49:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144315	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	250477	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104248	51.25	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.50%	
35) Dibromofluoromethane	5.391	113	83603	50.24	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.48%	
50) Toluene-d8	8.653	98	310759	50.74	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.48%	
62) 4-Bromofluorobenzene	11.085	95	95413	39.62	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 79.24%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024575.D
Acq On : 05 Oct 2021 16:49
Operator : JC/MD
Sample : M4080-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P287-DITCH

Quant Time: Oct 06 06:49:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

