

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024584.D
 Acq On : 05 Oct 2021 20:21
 Operator : JC/MD
 Sample : M4080-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P287-DITCH

Quant Time: Oct 06 06:51:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251491	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107781	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96299	55.52	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.04%
35) Dibromofluoromethane	5.391	113	73408	43.94	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.88%
50) Toluene-d8	8.653	98	313752	51.02	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.04%
62) 4-Bromofluorobenzene	11.085	95	115322	47.70	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024584.D
Acq On : 05 Oct 2021 20:21
Operator : JC/MD
Sample : M4080-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P287-DITCH

Quant Time: Oct 06 06:51:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

