

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024599.D
 Acq On : 06 Oct 2021 03:22
 Operator : JC/MD
 Sample : M4055-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-RW8-MW01D1-GW-20210930

Quant Time: Oct 06 06:53:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	146199	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255909	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104250	50.59	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.18%
35) Dibromofluoromethane	5.391	113	86041	50.61	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.22%
50) Toluene-d8	8.653	98	318008	50.82	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.64%
62) 4-Bromofluorobenzene	11.085	95	105974	43.08	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.16%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024599.D
Acq On : 06 Oct 2021 03:22
Operator : JC/MD
Sample : M4055-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-RW8-MW01D1-GW-20210930

Quant Time: Oct 06 06:53:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

