

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024602.D
 Acq On : 06 Oct 2021 04:32
 Operator : JC/MD
 Sample : M3960-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAMB-MW06-GW

Quant Time: Oct 06 06:54:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	249347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245814	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97729	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108397	51.53	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.06%			
35) Dibromofluoromethane	5.397	113	88265	53.28	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.56%			
50) Toluene-d8	8.653	98	328590	53.90	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 107.80%			
62) 4-Bromofluorobenzene	11.085	95	102427	42.73	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 85.46%			
Target Compounds					Qvalue	
31) Cyclohexane	5.476	56	3129	1.24	ug/l	93
39) Methylcyclohexane	7.379	83	3543	1.27	ug/l #	79
40) Benzene	6.050	78	12086	1.77	ug/l	98
73) Isopropylbenzene	10.963	105	37719	5.46	ug/l	98
78) n-propylbenzene	11.305	91	25144	3.11	ug/l	96
85) sec-Butylbenzene	11.890	105	9854	1.35	ug/l #	57
95) Naphthalene	13.780	128	38237	5.61	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024602.D
Acq On : 06 Oct 2021 04:32
Operator : JC/MD
Sample : M3960-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAMB-MW06-GW

Quant Time: Oct 06 06:54:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

