

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024603.D
 Acq On : 06 Oct 2021 04:56
 Operator : JC/MD
 Sample : M3960-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAMB-MW08-GW

Manual Integrations
 APPROVED

MMDadoda
 10/6/2021 7:03:59 PM

Quant Time: Oct 06 06:54:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	153025	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264256	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	224448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134946	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105286	48.82	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.64%		
35) Dibromofluoromethane	5.398	113	86700	49.39	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.78%		
50) Toluene-d8	8.653	98	279907	43.32	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	86.64%#		
62) 4-Bromofluorobenzene	11.086	95	131806	51.88	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.76%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	2.959	59	18131	25.02	ug/l #	73
16) Acetone	2.386	43	11110	2.16	ug/l	90
19) Methyl tert-butyl Ether	3.124	73	165142	31.24	ug/l #	61
25) 2-Butanone	4.574	43	8572	5.77	ug/l #	86
31) Cyclohexane	5.477	56	118799	46.00	ug/l	97
39) Methylcyclohexane	7.385	83	113339	38.21	ug/l	93
40) Benzene	6.044	78	711497	98.22	ug/l	98
52) Toluene	8.726	92	643151	137.44	ug/l	98
67) Ethyl Benzene	10.208	91	13669165	1651.48	ug/l #	65
68) m/p-Xylenes	10.330	106	16933857	5335.30	ug/l #	1
69) o-Xylene	10.665	106	11364622	3620.96	ug/l	62
73) Isopropylbenzene	10.970	105	1093680	114.64	ug/l	99
78) n-propylbenzene	11.311	91	2586233	231.90	ug/l	98
80) 1,3,5-Trimethylbenzene	11.463	105	5564769	682.81	ug/l	95
84) 1,2,4-Trimethylbenzene	11.768	105	15742264	1942.68	ug/l	73
85) sec-Butylbenzene	11.896	105	136109m	13.54	ug/l	
86) p-Isopropyltoluene	12.018	119	87821m	10.25	ug/l	
95) Naphthalene	13.786	128	9952049	1056.65	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024603.D
Acq On : 06 Oct 2021 04:56
Operator : JC/MD
Sample : M3960-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAMB-MW08-GW

Manual Integrations
APPROVED

MMDadoda
10/6/2021 7:03:59 PM

Quant Time: Oct 06 06:54:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

