

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024605.D
 Acq On : 06 Oct 2021 05:43
 Operator : JC/MD
 Sample : M3960-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAMB-MW11-GW

Quant Time: Oct 06 06:54:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	166150	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	293247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	254774	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123540	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113065	48.28	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.56%		
35) Dibromofluoromethane	5.397	113	95850	49.20	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.40%		
50) Toluene-d8	8.653	98	354455	49.44	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.88%		
62) 4-Bromofluorobenzene	11.085	95	134070	47.56	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.12%		
Target Compounds						Qvalue
30) Chloroform	5.111	83	1324	0.38	ug/l	89
67) Ethyl Benzene	10.195	91	7328	0.78	ug/l	93
68) m/p-Xylenes	10.305	106	9937	2.76	ug/l	97
69) o-Xylene	10.646	106	3690	1.04	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	2833	0.38	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	11517	1.55	ug/l	98
95) Naphthalene	13.780	128	4591	0.53	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024605.D
Acq On : 06 Oct 2021 05:43
Operator : JC/MD
Sample : M3960-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAMB-MW11-GW

Quant Time: Oct 06 06:54:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

