

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024608.D
 Acq On : 06 Oct 2021 06:53
 Operator : JC/MD
 Sample : M4088-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-20211005-WATER

Quant Time: Oct 06 09:10:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	161715	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247013	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107562	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111018	48.71	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.42%
35) Dibromofluoromethane	5.397	113	93487	56.97	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.94%
50) Toluene-d8	8.653	98	346165	57.32	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	114.64%#
62) 4-Bromofluorobenzene	11.085	95	112633	47.43	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.86%
Target Compounds						
					Qvalue	
25) 2-Butanone	4.574	43	4273	2.72	ug/l	96
30) Chloroform	5.105	83	37699	11.10	ug/l	100
47) Bromodichloromethane	7.824	83	1958	0.82	ug/l	92
52) Toluene	8.720	92	1633	0.37	ug/l	95
67) Ethyl Benzene	10.195	91	4126	0.48	ug/l	95
68) m/p-Xylenes	10.305	106	4570	1.39	ug/l	93
69) o-Xylene	10.646	106	1744	0.54	ug/l	90
80) 1,3,5-Trimethylbenzene	11.457	105	2030	0.31	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	9529	1.48	ug/l	96
95) Naphthalene	13.780	128	99996	13.32	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
Data File : VX024608.D
Acq On : 06 Oct 2021 06:53
Operator : JC/MD
Sample : M4088-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-20211005-WATER

Quant Time: Oct 06 09:10:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

