

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100521\  
 Data File : VX024565.D  
 Acq On : 05 Oct 2021 12:55  
 Operator : JC/MD  
 Sample : VX1005WBSD01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1005WBSD01

Manual Integrations  
 APPROVED

MMDadoda  
 10/6/2021 7:03:50 PM

Quant Time: Oct 05 16:41:39 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092321W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 23 18:24:34 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc   | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|--------|--------|----------|
| -----                        |                |      |            |        |        |          |
| Internal Standards           |                |      |            |        |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 180867     | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 286056     | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 250629     | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 127595     | 50.00  | ug/l   | 0.00     |
|                              |                |      |            |        |        |          |
| System Monitoring Compounds  |                |      |            |        |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964          | 65   | 123062     | 48.28  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = | 96.56% |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 94717      | 49.84  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 99.68% |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 348879     | 49.88  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = | 99.76% |        |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 128275     | 46.65  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = | 93.30% |        |          |
|                              |                |      |            |        |        |          |
| Target Compounds             |                |      |            |        | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.173          | 85   | 32474      | 17.84  | ug/l   | 97       |
| 3) Chloromethane             | 1.294          | 50   | 32130      | 18.73  | ug/l   | 100      |
| 4) Vinyl Chloride            | 1.374          | 62   | 33188      | 18.49  | ug/l   | 100      |
| 5) Bromomethane              | 1.599          | 94   | 20955      | 17.35  | ug/l   | 93       |
| 6) Chloroethane              | 1.685          | 64   | 18596      | 15.74  | ug/l   | 98       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 50342      | 15.90  | ug/l   | 99       |
| 8) Diethyl Ether             | 2.136          | 74   | 18536      | 17.20  | ug/l   | 91       |
| 9) 1,1,2-Trichlorotrifluo... | 2.331          | 101  | 34184      | 19.14  | ug/l   | 95       |
| 10) Methyl Iodide            | 2.453          | 142  | 36520      | 14.62  | ug/l   | 95       |
| 11) Tert butyl alcohol       | 2.971          | 59   | 52570      | 80.28  | ug/l   | 98       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 31218      | 18.83  | ug/l   | 93       |
| 13) Acrolein                 | 2.239          | 56   | 11987      | 93.51  | ug/l   | 99       |
| 14) Allyl chloride           | 2.666          | 41   | 57892      | 19.10  | ug/l   | 96       |
| 15) Acrylonitrile            | 3.068          | 53   | 116661     | 100.74 | ug/l   | 99       |
| 16) Acetone                  | 2.386          | 43   | 106069     | 92.22  | ug/l   | 93       |
| 17) Carbon Disulfide         | 2.514          | 76   | 71304      | 16.84  | ug/l   | 98       |
| 18) Methyl Acetate           | 2.709          | 43   | 75231      | 19.71  | ug/l   | 92       |
| 19) Methyl tert-butyl Ether  | 3.123          | 73   | 125735     | 20.12  | ug/l   | 95       |
| 20) Methylene Chloride       | 2.794          | 84   | 39097      | 19.81  | ug/l   | 93       |
| 21) trans-1,2-Dichloroethene | 3.099          | 96   | 34741      | 18.45  | ug/l   | 99       |
| 22) Diisopropyl ether        | 3.770          | 45   | 120491     | 19.46  | ug/l   | 97       |
| 23) Vinyl Acetate            | 3.727          | 43   | 491090     | 97.93  | ug/l   | 95       |
| 24) 1,1-Dichloroethane       | 3.617          | 63   | 67330      | 19.58  | ug/l   | 98       |
| 25) 2-Butanone               | 4.568          | 43   | 167315     | 95.31  | ug/l   | 93       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 55738      | 18.02  | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.495          | 96   | 42442      | 19.29  | ug/l   | 99       |
| 28) Bromochloromethane       | 4.910          | 49   | 26001      | 16.66  | ug/l   | 98       |
| 29) Tetrahydrofuran          | 5.019          | 42   | 107244     | 95.87  | ug/l   | 91       |
| 30) Chloroform               | 5.105          | 83   | 72337      | 19.05  | ug/l   | 96       |
| 31) Cyclohexane              | 5.477          | 56   | 56421      | 18.48  | ug/l   | 97       |
| 32) 1,1,1-Trichloroethane    | 5.391          | 97   | 62384      | 18.50  | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 5.702          | 75   | 48917      | 17.84  | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.727          | 43   | 62787      | 18.34  | ug/l   | 95       |
| 38) Carbon Tetrachloride     | 5.684          | 117  | 53835      | 18.03  | ug/l   | 100      |
| 39) Methylcyclohexane        | 7.385          | 83   | 58656      | 18.27  | ug/l   | 97       |
| 40) Benzene                  | 6.044          | 78   | 148459     | 18.93  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100521\  
 Data File : VX024565.D  
 Acq On : 05 Oct 2021 12:55  
 Operator : JC/MD  
 Sample : VX1005WBSD01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1005WBSD01

Manual Integrations  
 APPROVED

MMDadoda  
 10/6/2021 7:03:50 PM

Quant Time: Oct 05 16:41:39 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092321W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 23 18:24:34 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 41) Methacrylonitrile         | 4.934  | 41   | 34278    | 19.39  | ug/l   | 94       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 59137    | 17.96  | ug/l   | 96       |
| 43) Isopropyl Acetate         | 6.348  | 43   | 95350    | 18.32  | ug/l   | 95       |
| 44) Trichloroethene           | 7.129  | 130  | 40509    | 18.78  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 7.434  | 63   | 39790    | 19.67  | ug/l   | 99       |
| 46) Dibromomethane            | 7.586  | 93   | 28178    | 18.57  | ug/l   | 98       |
| 47) Bromodichloromethane      | 7.830  | 83   | 51978    | 18.70  | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.696  | 41   | 47602    | 18.47  | ug/l   | 90       |
| 49) 1,4-Dioxane               | 7.659  | 88   | 21720    | 367.92 | ug/l   | 98       |
| 51) 4-Methyl-2-Pentanone      | 8.580  | 43   | 323886   | 94.25  | ug/l   | 93       |
| 52) Toluene                   | 8.720  | 92   | 96526    | 19.06  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 58195    | 18.47  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 8.372  | 75   | 62416    | 19.48  | ug/l   | 99       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 42399    | 20.24  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 64398    | 19.32  | ug/l   | 92       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 69607    | 19.70  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 144528   | 87.83  | ug/l   | 98       |
| 59) 2-Hexanone                | 9.433  | 43   | 240723   | 88.95  | ug/l   | 89       |
| 60) Dibromochloromethane      | 9.525  | 129  | 39005    | 17.21  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.616  | 107  | 42086    | 18.30  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 35707    | 20.59  | ug/l   | 93       |
| 65) Chlorobenzene             | 10.079 | 112  | 99988    | 19.15  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 38070    | 19.56  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 172586   | 18.67  | ug/l   | 97       |
| 68) m/p-Xylenes               | 10.305 | 106  | 135926   | 38.35  | ug/l   | 99       |
| 69) o-Xylene                  | 10.646 | 106  | 66737    | 19.04  | ug/l   | 98       |
| 70) Styrene                   | 10.659 | 104  | 112533   | 19.20  | ug/l   | 97       |
| 71) Bromoform                 | 10.805 | 173  | 28003    | 18.07  | ug/l # | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 181251   | 20.09  | ug/l   | 100      |
| 74) N-amyl acetate            | 10.848 | 43   | 78688    | 18.97  | ug/l   | 93       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 63699    | 19.97  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 60332m   | 20.82  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 44978    | 19.38  | ug/l   | 97       |
| 78) n-propylbenzene           | 11.305 | 91   | 201379   | 19.10  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 125224   | 19.49  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 153295   | 19.89  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 17961    | 18.36  | ug/l   | 86       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 148110   | 19.60  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 150889   | 19.53  | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 156672   | 20.45  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 188810   | 19.87  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 157417   | 19.42  | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 83044    | 18.79  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 83087    | 18.20  | ug/l   | 98       |
| 89) n-Butylbenzene            | 12.335 | 91   | 136459   | 19.02  | ug/l   | 97       |
| 90) Hexachloroethane          | 12.542 | 117  | 24054    | 18.05  | ug/l   | 95       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 83496    | 19.11  | ug/l   | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 14653    | 19.02  | ug/l   | 90       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 51603    | 18.55  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 23046    | 18.11  | ug/l   | 100      |
| 95) Naphthalene               | 13.780 | 128  | 170620   | 19.16  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 53311    | 18.70  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100521\  
Data File : VX024565.D  
Acq On : 05 Oct 2021 12:55  
Operator : JC/MD  
Sample : VX1005WBSD01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM

Quant Time: Oct 05 16:41:39 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092321W.M  
Quant Title : SW846 8260  
QLast Update : Thu Sep 23 18:24:34 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100521\  
Data File : VX024565.D  
Acq On : 05 Oct 2021 12:55  
Operator : JC/MD  
Sample : VX1005WBSD01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 6 Sample Multiplier: 1


Instrument :  
MSVOA\_X  
Client Sampled :  
VX1005WBSD01

Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM

Quant Time: Oct 05 16:41:39 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092321W.M  
Quant Title : SW846 8260  
QLast Update : Thu Sep 23 18:24:34 2021  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.00 ug/l  
RT: 5.556 min Scan# 733  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion: 168 Resp: 180867  
Ion Ratio Lower Upper  
168 100  
99 58.0 45.8 68.6



#2  
Dichlorodifluoromethane  
Concen: 17.84 ug/l  
RT: 1.173 min Scan# 14  
Delta R.T. 0.007 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55



Tgt Ion: 85 Resp: 32474  
Ion Ratio Lower Upper  
85 100  
87 31.4 16.4 49.2




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM

Abundance Scan 34 (1.294 min): VX024453.D\data.ms (-30)







#3  
Chloromethane  
Concen: 18.73 ug/l  
RT: 1.294 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 50 Resp: 32130  
Ion Ratio Lower Upper  
50 100  
52 32.9 26.4 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#4  
Vinyl Chloride  
Concen: 18.49 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 62 Resp: 33188  
Ion Ratio Lower Upper  
62 100  
64 31.9 25.7 38.5









#5  
Bromomethane  
Concen: 17.35 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 94 Resp: 20955  
Ion Ratio Lower Upper  
94 100  
96 88.3 76.4 114.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#6  
Chloroethane  
Concen: 15.74 ug/l  
RT: 1.685 min Scan# 98  
Delta R.T. 0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 64 Resp: 18596  
Ion Ratio Lower Upper  
64 100  
66 32.6 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 15.90 ug/l  
RT: 1.886 min Scan# 131  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion:101 Resp: 50342  
Ion Ratio Lower Upper  
101 100  
103 67.3 53.2 79.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM




#8  
Diethyl Ether  
Concen: 17.20 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 74 Resp: 18536  
Ion Ratio Lower Upper  
74 100  
45 97.0 44.5 133.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 19.14 ug/l  
RT: 2.331 min Scan# 204  
Delta R.T. 0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion:101 Resp: 34184  
Ion Ratio Lower Upper  
101 100  
85 45.2 34.9 52.3  
151 82.0 61.4 92.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#10  
Methyl Iodide  
Concen: 14.62 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion:142 Resp: 36520  
Ion Ratio Lower Upper  
142 100  
127 45.9 33.3 49.9  
141 14.6 11.4 17.0









#11  
Tert butyl alcohol  
Concen: 80.28 ug/l  
RT: 2.971 min Scan# 309  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 59 Resp: 52570  
Ion Ratio Lower Upper  
59 100  
57 10.4 7.8 11.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#12  
1,1-Dichloroethene  
Concen: 18.83 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 96 Resp: 31218  
Ion Ratio Lower Upper  
96 100  
61 168.3 124.6 187.0  
98 62.6 50.6 76.0









#13  
Acrolein  
Concen: 93.51 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 56 Resp: 11987  
Ion Ratio Lower Upper  
56 100  
55 69.3 56.4 84.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM




#14  
Allyl chloride  
Concen: 19.10 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 41 Resp: 57892  
Ion Ratio Lower Upper  
41 100  
39 63.3 51.6 77.4  
76 32.6 30.5 45.7









#15  
Acrylonitrile  
Concen: 100.74 ug/l  
RT: 3.068 min Scan# 325  
Delta R.T. -0.001 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion: 53 Resp: 116661  
Ion Ratio Lower Upper  
53 100  
52 82.4 65.3 97.9  
51 36.8 28.2 42.2



#16  
Acetone  
Concen: 92.22 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 43 Resp: 106069  
Ion Ratio Lower Upper  
43 100  
58 30.6 27.8 41.8



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM







#17  
Carbon Disulfide  
Concen: 16.84 ug/l  
RT: 2.514 min Scan# 234  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 76 Resp: 71304  
Ion Ratio Lower Upper  
76 100  
78 9.6 7.2 10.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#18  
Methyl Acetate  
Concen: 19.71 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 75231  
Ion Ratio Lower Upper  
43 100  
74 23.4 21.9 32.9









#19  
Methyl tert-butyl Ether  
Concen: 20.12 ug/l  
RT: 3.123 min Scan# 334  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 73 Resp: 125735  
Ion Ratio Lower Upper  
73 100  
57 19.2 17.1 25.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM




#20  
Methylene Chloride  
Concen: 19.81 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 84 Resp: 39097  
Ion Ratio Lower Upper  
84 100  
49 119.1 88.6 132.8  
51 39.3 27.0 40.6  
86 66.6 50.3 75.5









#21  
trans-1,2-Dichloroethene  
Concen: 18.45 ug/l  
RT: 3.099 min Scan# 330  
Delta R.T. 0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion: 96 Resp: 34741  
Ion Ratio Lower Upper  
96 100  
61 133.8 107.7 161.5  
98 62.6 50.4 75.6



#22  
Diisopropyl ether  
Concen: 19.46 ug/l  
RT: 3.770 min Scan# 440  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 45 Resp: 120491  
Ion Ratio Lower Upper  
45 100  
43 74.9 58.0 87.0  
87 29.6 25.4 38.0  
59 13.4 10.2 15.2



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM







#23  
 Vinyl Acetate  
 Concen: 97.93 ug/l  
 RT: 3.727 min Scan# 433  
 Delta R.T. -0.006 min  
 Lab File: VX024565.D  
 Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 491090  
 Ion Ratio Lower Upper  
 43 100  
 86 10.4 9.9 14.9

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1005WBSD01




Manual Integrations  
 APPROVED

MMDadoda  
 10/6/2021 7:03:50 PM




#24  
 1,1-Dichloroethane  
 Concen: 19.58 ug/l  
 RT: 3.617 min Scan# 415  
 Delta R.T. 0.000 min  
 Lab File: VX024565.D  
 Acq: 05 Oct 2021 12:55

Tgt Ion: 63 Resp: 67330  
 Ion Ratio Lower Upper  
 63 100  
 98 6.9 3.8 11.2  
 100 4.5 2.5 7.5











#25  
2-Butanone  
Concen: 95.31 ug/l  
RT: 4.568 min Scan# 571  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 167315  
Ion Ratio Lower Upper  
43 100  
72 25.7 23.7 35.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#26  
2,2-Dichloropropane  
Concen: 18.02 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 77 Resp: 55738  
Ion Ratio Lower Upper  
77 100  
97 22.5 11.7 35.1









#27  
cis-1,2-Dichloroethene  
Concen: 19.29 ug/l  
RT: 4.495 min Scan# 559  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

|             |             |       |       |
|-------------|-------------|-------|-------|
| Tgt Ion: 96 | Resp: 42442 |       |       |
| Ion Ratio   | Lower       | Upper |       |
| 96          | 100         |       |       |
| 61          | 143.6       | 0.0   | 285.6 |
| 98          | 63.9        | 0.0   | 129.6 |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#28  
Bromochloromethane  
Concen: 16.66 ug/l  
RT: 4.910 min Scan# 627  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55


|             |             |
|-------------|-------------|
| Tgt Ion: 49 | Resp: 26001 |
| Ion Ratio   | Lower Upper |
| 49 100      |             |
| 129 2.3     | 0.0 4.0     |
| 130 83.7    | 68.2 102.2  |





#29  
Tetrahydrofuran  
Concen: 95.87 ug/l  
RT: 5.019 min Scan# 645  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Tgt Ion: 42 Resp: 107244  
Ion Ratio Lower Upper  
42 100  
72 46.0 41.7 62.5  
71 42.2 38.4 57.6

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#30  
Chloroform  
Concen: 19.05 ug/l  
RT: 5.105 min Scan# 659  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 83 Resp: 72337  
Ion Ratio Lower Upper  
83 100  
85 68.5 52.0 78.0









#31  
Cyclohexane  
Concen: 18.48 ug/l  
RT: 5.477 min Scan# 720  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion: 56 Resp: 56421  
Ion Ratio Lower Upper  
56 100  
69 30.4 25.3 37.9  
84 85.6 71.4 107.0



#32  
1,1,1-Trichloroethane  
Concen: 18.50 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 97 Resp: 62384  
Ion Ratio Lower Upper  
97 100  
99 65.1 51.5 77.3  
61 41.2 34.2 51.4



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM







#33  
 1,2-Dichloroethane-d4  
 Concen: 48.28 ug/l  
 RT: 5.964 min Scan# 800  
 Delta R.T. 0.000 min  
 Lab File: VX024565.D  
 Acq: 05 Oct 2021 12:55

Tgt Ion: 65 Resp: 123062  
 Ion Ratio Lower Upper  
 65 100  
 67 51.6 0.0 105.4

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1005WBSD01




Manual Integrations  
 APPROVED


MMDadoda  
 10/6/2021 7:03:50 PM



#34  
 1,4-Difluorobenzene  
 Concen: 50.00 ug/l  
 RT: 6.763 min Scan# 931  
 Delta R.T. -0.006 min  
 Lab File: VX024565.D  
 Acq: 05 Oct 2021 12:55

Tgt Ion: 114 Resp: 286056  
 Ion Ratio Lower Upper  
 114 100  
 63 20.7 0.0 40.8  
 88 15.6 0.0 32.8









#35  
Dibromofluoromethane  
Concen: 49.84 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 113 Resp: 94717  
Ion Ratio Lower Upper  
113 100  
111 103.2 81.9 122.9  
192 20.2 16.7 25.1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#36  
1,1-Dichloropropene  
Concen: 17.84 ug/l  
RT: 5.702 min Scan# 757  
Delta R.T. 0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 75 Resp: 48917  
Ion Ratio Lower Upper  
75 100  
110 36.4 18.1 54.4  
77 31.0 25.3 37.9





#37

Ethyl Acetate

Concen: 18.34 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.000 min

Lab File: VX024565.D





Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 43    | Resp: | 62787 |
| Ion      | Ratio | Lower | Upper |
| 43       | 100   |       |       |
| 61       | 14.4  | 12.8  | 19.2  |
| 70       | 10.8  | 10.5  | 15.7  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM



#38

Carbon Tetrachloride

Concen: 18.03 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.006 min

Lab File: VX024565.D

Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 117   | Resp: | 53835 |
| Ion      | Ratio | Lower | Upper |
| 117      | 100   |       |       |
| 119      | 95.6  | 76.2  | 114.2 |
| 121      | 31.3  | 24.8  | 37.2  |





#39

Methylcyclohexane

Concen: 18.27 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX024565.D

Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 58656 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 55       | 77.3  | 58.9  | 88.3  |
| 98       | 46.3  | 38.1  | 57.1  |

Instrument :





MSVOA\_X

ClientSampleId :

VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM



#40

Benzene

Concen: 18.93 ug/l




RT: 6.044 min Scan# 813

Delta R.T. -0.000 min


Lab File: VX024565.D

Acq: 05 Oct 2021 12:55

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 78    | Resp: | 148459 |
| Ion      | Ratio | Lower | Upper  |
| 78       | 100   |       |        |
| 77       | 22.5  | 18.9  | 28.3   |











#41  
Methacrylonitrile  
Concen: 19.39 ug/l  
RT: 4.934 min Scan# 631  
Delta R.T. 0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 34278 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 54.9  | 42.5  | 63.7  |
| 67       | 73.9  | 65.0  | 97.6  |
| 52       | 29.5  | 25.3  | 37.9  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#42  
1,2-Dichloroethane  
Concen: 17.96 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 62    | Resp: | 59137 |
| Ion      | Ratio | Lower | Upper |
| 62       | 100   |       |       |
| 98       | 8.5   | 0.0   | 20.2  |









#43  
Isopropyl Acetate  
Concen: 18.32 ug/l  
RT: 6.348 min Scan# 863  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 95350  
Ion Ratio Lower Upper  
43 100  
61 20.3 18.2 27.4  
87 13.7 12.5 18.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#44  
Trichloroethene  
Concen: 18.78 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion:130 Resp: 40509  
Ion Ratio Lower Upper  
130 100  
95 99.2 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 19.67 ug/l  
RT: 7.434 min Scan# 1041  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 63 Resp: 39790  
Ion Ratio Lower Upper  
63 100  
65 31.4 25.7 38.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED




MMDadoda  
10/6/2021 7:03:50 PM




#46

Dibromomethane  
Concen: 18.57 ug/l  
RT: 7.586 min Scan# 1066  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 93 Resp: 28178  
Ion Ratio Lower Upper  
93 100  
95 82.3 66.8 100.2  
174 102.0 83.1 124.7









#47  
Bromodichloromethane  
Concen: 18.70 ug/l  
RT: 7.830 min Scan# 1106  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 51978 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 85       | 65.4  | 51.4  | 77.0  |
| 127      | 9.0   | 6.8   | 10.2  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#48  
Methyl methacrylate  
Concen: 18.47 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 47602 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 69       | 81.5  | 76.0  | 114.0 |
| 39       | 54.1  | 44.6  | 67.0  |









#49  
1,4-Dioxane  
Concen: 367.92 ug/l  
RT: 7.659 min Scan# 1078  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 88 Resp: 21720  
Ion Ratio Lower Upper  
88 100  
43 30.6 23.2 34.8  
58 69.7 55.0 82.4

Instrument :  
MSVOA\_X  
ClientSampled :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#50  
Toluene-d8  
Concen: 49.88 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 98 Resp: 348879  
Ion Ratio Lower Upper  
98 100  
100 64.6 51.9 77.9









#51  
4-Methyl-2-Pentanone  
Concen: 94.25 ug/l  
RT: 8.580 min Scan# 1229  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 323886  
Ion Ratio Lower Upper  
43 100  
58 39.5 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#52  
Toluene  
Concen: 19.06 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 92 Resp: 96526  
Ion Ratio Lower Upper  
92 100  
91 173.7 137.1 205.7









#53  
t-1,3-Dichloropropene  
Concen: 18.47 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 75 Resp: 58195  
Ion Ratio Lower Upper  
75 100  
77 32.5 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM




#54  
cis-1,3-Dichloropropene  
Concen: 19.48 ug/l  
RT: 8.372 min Scan# 1195  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 75 Resp: 62416  
Ion Ratio Lower Upper  
75 100  
77 32.7 25.7 38.5  
39 44.4 35.1 52.7








#55  
1,1,2-Trichloroethane  
Concen: 20.24 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 97 Resp: 42399

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 83.3  | 69.3  | 103.9 |
| 85  | 55.3  | 44.3  | 66.5  |
| 99  | 63.3  | 50.3  | 75.5  |





#56  
Ethyl methacrylate  
Concen: 19.32 ug/l  
RT: 9.122 min Scan# 1318  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55



Tgt Ion: 69 Resp: 64398

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 64.3  | 46.8  | 70.2  |
| 39  | 34.8  | 24.7  | 37.1  |




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM









#57  
1,3-Dichloropropane  
Concen: 19.70 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 76 Resp: 69607  
Ion Ratio Lower Upper  
76 100  
78 31.7 26.2 39.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#58  
2-Chloroethyl Vinyl ether  
Concen: 87.83 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 63 Resp: 144528  
Ion Ratio Lower Upper  
63 100  
106 28.4 23.5 35.3









#59  
2-Hexanone  
Concen: 88.95 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 240723  
Ion Ratio Lower Upper  
43 100  
58 53.1 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#60  
Dibromochloromethane  
Concen: 17.21 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55


Tgt Ion: 129 Resp: 39005  
Ion Ratio Lower Upper  
129 100  
127 76.8 38.4 115.2





#61  
1,2-Dibromoethane  
Concen: 18.30 ug/l  
RT: 9.616 min Scan# 1399  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Tgt Ion:107 Resp: 42086  
Ion Ratio Lower Upper  
107 100  
109 93.9 75.8 113.6

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#62  
4-Bromofluorobenzene  
Concen: 46.65 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 95 Resp: 128275  
Ion Ratio Lower Upper  
95 100  
174 80.1 0.0 154.6  
176 77.0 0.0 155.8







#63  
Chlorobenzene-d5  
Concen: 50.00 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion:117 Resp: 250629  
Ion Ratio Lower Upper  
117 100  
82 56.3 47.4 71.2  
119 30.8 26.6 39.8




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#64  
Tetrachloroethene  
Concen: 20.59 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55



Tgt Ion:164 Resp: 35707  
Ion Ratio Lower Upper  
164 100  
166 134.5 103.4 155.2  
129 100.5 70.8 106.2  
131 96.0 72.5 108.7









#65  
Chlorobenzene  
Concen: 19.15 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. -0.007 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Tgt Ion:112 Resp: 99988  
Ion Ratio Lower Upper  
112 100  
114 33.3 25.8 38.6




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.56 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion:131 Resp: 38070  
Ion Ratio Lower Upper  
131 100  
133 94.3 47.7 143.1  
119 64.3 32.1 96.5









#67  
Ethyl Benzene  
Concen: 18.67 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion: 91 Resp: 172586  
Ion Ratio Lower Upper  
91 100  
106 32.4 24.6 37.0



#68  
m/p-Xylenes  
Concen: 38.35 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 106 Resp: 135926  
Ion Ratio Lower Upper  
106 100  
91 205.6 162.8 244.2



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01


Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM



#69  
o-Xylene  
Concen: 19.04 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Tgt Ion:106 Resp: 66737  
Ion Ratio Lower Upper  
106 100  
91 218.9 108.1 324.2

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#70  
Styrene  
Concen: 19.20 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion:104 Resp: 112533  
Ion Ratio Lower Upper  
104 100  
78 52.7 40.0 60.0  
103 53.3 43.4 65.0









#71  
Bromoform  
Concen: 18.07 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion:173 Resp: 28003  
Ion Ratio Lower Upper  
173 100  
175 48.5 23.9 71.8  
254 0.0 0.1 0.1#



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.00 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55



Tgt Ion:152 Resp: 127595  
Ion Ratio Lower Upper  
152 100  
115 70.3 41.1 123.3  
150 164.8 0.0 349.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM









#73  
Isopropylbenzene  
Concen: 20.09 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 105 Resp: 181251  
Ion Ratio Lower Upper  
105 100  
120 26.5 13.2 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#74  
N-ethyl acetate  
Concen: 18.97 ug/l  
RT: 10.848 min Scan# 1601  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 43 Resp: 78688  
Ion Ratio Lower Upper  
43 100  
70 42.6 39.0 58.6  
55 25.0 21.5 32.3  
61 23.6 22.1 33.1









#75  
1,1,2,2-Tetrachloroethane  
Concen: 19.97 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 83 Resp: 63699  
Ion Ratio Lower Upper  
83 100  
131 11.5 5.1 15.2  
85 63.5 31.8 95.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#76  
1,2,3-Trichloropropane  
Concen: 20.82 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 75 Resp: 60332  
Ion Ratio Lower Upper  
75 100  
77 38.7 19.1 57.5





#77  
Bromobenzene  
Concen: 19.38 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion: 156 Resp: 44978  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 156 | 100   |       |       |
| 77  | 158.9 | 82.8  | 248.6 |
| 158 | 98.0  | 49.1  | 147.4 |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED




MMDadoda  
10/6/2021 7:03:50 PM




#78  
n-propylbenzene  
Concen: 19.10 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 91 Resp: 201379  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 91  | 100   |       |       |
| 120 | 23.6  | 11.5  | 34.5  |









#79  
2-Chlorotoluene  
Concen: 19.49 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Tgt Ion: 91 Resp: 125224  
Ion Ratio Lower Upper  
91 100  
126 34.4 16.6 49.8




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 19.89 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55


Tgt Ion: 105 Resp: 153295  
Ion Ratio Lower Upper  
105 100  
120 47.3 23.8 71.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 18.36 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. -0.006 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Tgt Ion: 75 Resp: 17961  
Ion Ratio Lower Upper  
75 100  
53 111.0 74.5 111.7  
89 49.7 37.9 56.9

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#82  
4-Chlorotoluene  
Concen: 19.60 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 91 Resp: 148110  
Ion Ratio Lower Upper  
91 100  
126 29.4 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 19.53 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Tgt Ion:119 Resp: 150889  
Ion Ratio Lower Upper  
119 100  
91 59.6 27.9 83.7  
134 22.5 11.3 33.9

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#84  
1,2,4-Trimethylbenzene  
Concen: 20.45 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion:105 Resp: 156672  
Ion Ratio Lower Upper  
105 100  
120 44.0 22.4 67.3





#85  
 sec-Butylbenzene  
 Concen: 19.87 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. -0.006 min  
 Lab File: VX024565.D  
 Acq: 05 Oct 2021 12:55




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1005WBSD01




Tgt Ion:105 Resp: 188810  
 Ion Ratio Lower Upper  
 105 100  
 134 19.9 10.0 29.8

Manual Integrations  
 APPROVED



MMDadoda  
 10/6/2021 7:03:50 PM




#86  
 p-Isopropyltoluene  
 Concen: 19.42 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. 0.000 min  
 Lab File: VX024565.D  
 Acq: 05 Oct 2021 12:55



Tgt Ion:119 Resp: 157417  
 Ion Ratio Lower Upper  
 119 100  
 134 26.2 12.9 38.7  
 91 25.2 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 18.79 ug/l  
RT: 11.969 min Scan# 1785  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion:146 Resp: 83044  
Ion Ratio Lower Upper  
146 100  
111 41.2 20.2 60.5  
148 63.9 31.9 95.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM




#88  
1,4-Dichlorobenzene  
Concen: 18.20 ug/l  
RT: 12.043 min Scan# 1797  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55


Tgt Ion:146 Resp: 83087  
Ion Ratio Lower Upper  
146 100  
111 42.0 19.6 58.8  
148 63.3 31.6 95.0









#89  
n-Butylbenzene  
Concen: 19.02 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion: 91 Resp: 136459  
Ion Ratio Lower Upper  
91 100  
92 52.3 27.4 82.0  
134 25.0 13.2 39.5




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM




#90  
Hexachloroethane  
Concen: 18.05 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55



Tgt Ion: 117 Resp: 24054  
Ion Ratio Lower Upper  
117 100  
201 76.2 40.4 121.2









#91  
1,2-Dichlorobenzene  
Concen: 19.11 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 146 Resp: 83496  
Ion Ratio Lower Upper  
146 100  
111 42.0 20.8 62.2  
148 63.6 31.9 95.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01




Manual Integrations  
APPROVED


MMDadoda  
10/6/2021 7:03:50 PM




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 19.02 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55

Tgt Ion: 75 Resp: 14653  
Ion Ratio Lower Upper  
75 100  
155 79.5 45.3 135.9  
157 105.6 57.7 173.1









#93  
1,2,4-Trichlorobenzene  
Concen: 18.55 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55





Tgt Ion:180 Resp: 51603  
Ion Ratio Lower Upper  
180 100  
182 93.0 47.7 143.1  
145 30.7 15.0 45.0



#94  
Hexachlorobutadiene  
Concen: 18.11 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion:225 Resp: 23046  
Ion Ratio Lower Upper  
225 100  
223 62.0 31.1 93.2  
227 62.8 31.6 95.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED



MMDadoda  
10/6/2021 7:03:50 PM



#95  
Naphthalene  
Concen: 19.16 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. 0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55




Tgt Ion:128 Resp: 170620  
Ion Ratio Lower Upper  
128 100  
127 12.9 10.3 15.5  
129 11.2 8.8 13.2




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1005WBSD01

Manual Integrations  
APPROVED

MMDadoda  
10/6/2021 7:03:50 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 18.70 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX024565.D  
Acq: 05 Oct 2021 12:55



Tgt Ion:180 Resp: 53311  
Ion Ratio Lower Upper  
180 100  
182 93.7 47.6 142.9  
145 34.2 16.1 48.3

