



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100521\
 Data File : VX024560.D
 Acq On : 05 Oct 2021 09:21
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Title : SW846 8260
 Last Update : Thu Sep 23 18:24:34 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	4529	PASS
75	95	30	60	52.3	12145	PASS
95	95	100	100	100.0	23211	PASS
96	95	5	9	7.3	1688	PASS
173	174	0.00	2	0.9	173	PASS
174	95	50	100	80.4	18673	PASS
175	174	5	9	6.9	1283	PASS
176	174	95	101	97.9	18289	PASS
177	176	5	9	6.5	1182	PASS