

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
 Data File : VX031902.D
 Acq On : 05 Oct 2022 18:02
 Operator : JC/MD
 Sample : N5005-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-02-255-275-100422

Quant Time: Oct 06 06:47:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 05 12:20:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84665	51.904	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.800%			
35) Dibromofluoromethane	5.385	113	82829	48.113	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.220%			
50) Toluene-d8	8.653	98	299736	49.152	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.300%			
62) 4-Bromofluorobenzene	11.079	95	103278	47.083	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.160%			
Target Compounds					Qvalue	
16) Acetone	2.379	43	2592	5.486	ug/l	99
52) Toluene	8.720	92	22805	5.360	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
Data File : VX031902.D
Acq On : 05 Oct 2022 18:02
Operator : JC/MD
Sample : N5005-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-02-255-275-100422

Quant Time: Oct 06 06:47:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 05 12:20:53 2022
Response via : Initial Calibration

