

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
 Data File : VX031903.D
 Acq On : 05 Oct 2022 18:26
 Operator : JC/MD
 Sample : N5005-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-02-334-354-100422

Quant Time: Oct 06 06:48:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 05 12:20:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160706	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83482	52.917	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.840%
35) Dibromofluoromethane	5.385	113	79418	48.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.647	98	291340	49.754	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.500%
62) 4-Bromofluorobenzene	11.079	95	99780	47.372	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.740%
Target Compounds						
52) Toluene	8.720	92	36734	8.992	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
Data File : VX031903.D
Acq On : 05 Oct 2022 18:26
Operator : JC/MD
Sample : N5005-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-02-334-354-100422

Quant Time: Oct 06 06:48:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 05 12:20:53 2022
Response via : Initial Calibration

