

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
 Data File : VX031904.D
 Acq On : 05 Oct 2022 18:49
 Operator : JC/MD
 Sample : N4983-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-743-745

Quant Time: Oct 06 06:48:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 05 12:20:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158023	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112928	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84786	51.720	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.440%			
35) Dibromofluoromethane	5.385	113	82501	47.718	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.440%			
50) Toluene-d8	8.647	98	303956	49.632	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.260%			
62) 4-Bromofluorobenzene	11.079	95	104641	47.501	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.000%			
Target Compounds						Qvalue
7) Trichlorofluoromethane	1.886	101	1095	0.367	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.331	101	94868	65.694	ug/l #	84
12) 1,1-Dichloroethene	2.319	96	36754	26.875	ug/l #	77
16) Acetone	2.386	43	1952	4.111	ug/l #	82
24) 1,1-Dichloroethane	3.611	63	9966	4.377	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	7245	4.039	ug/l	70
30) Chloroform	5.093	83	3563	1.276	ug/l	95
32) 1,1,1-Trichloroethane	5.385	97	4417	1.797	ug/l	95
38) Carbon Tetrachloride	5.684	117	3948	1.671	ug/l	88
44) Trichloroethene	7.129	130	3931973	1914.540	ug/l	83
55) 1,1,2-Trichloroethane	9.153	97	3096	1.762	ug/l	98
64) Tetrachloroethene	9.275	164	8206	3.977	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
Data File : VX031904.D
Acq On : 05 Oct 2022 18:49
Operator : JC/MD
Sample : N4983-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-743-745

Quant Time: Oct 06 06:48:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 05 12:20:53 2022
Response via : Initial Calibration

