

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
 Data File : VX031905.D
 Acq On : 05 Oct 2022 19:12
 Operator : JC/MD
 Sample : N4983-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-763-765

Quant Time: Oct 06 06:48:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 05 12:20:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	153912	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162817	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84692	53.171	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.340%	
35) Dibromofluoromethane	5.385	113	81237	48.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.480%	
50) Toluene-d8	8.653	98	299570	50.222	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	11.079	95	101913	47.498	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.000%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	44148	31.464	ug/l #	85
12) 1,1-Dichloroethene	2.319	96	10762	8.099	ug/l #	78
16) Acetone	2.386	43	1358	2.943	ug/l #	86
27) cis-1,2-Dichloroethene	4.495	96	3386	1.943	ug/l	67
30) Chloroform	5.105	83	2103	0.775	ug/l #	68
44) Trichloroethene	7.129	130	1252298	626.050	ug/l	81
55) 1,1,2-Trichloroethane	9.159	97	1375	0.803	ug/l	90
64) Tetrachloroethene	9.281	164	2341	1.169	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100522\
Data File : VX031905.D
Acq On : 05 Oct 2022 19:12
Operator : JC/MD
Sample : N4983-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-763-765

Quant Time: Oct 06 06:48:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100522W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 05 12:20:53 2022
Response via : Initial Calibration

