

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100523\
 Data File : VX038098.D
 Acq On : 05 Oct 2023 13:27
 Operator : JC/MD
 Sample : VX1005MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1005MBL01

Quant Time: Oct 05 15:35:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250303	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	102473	44.181	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.360%
35) Dibromofluoromethane	5.385	113	81207	45.025	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	8.647	98	298628	43.330	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	86.660%#
62) 4-Bromofluorobenzene	11.079	95	120562	46.143	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100523\
Data File : VX038098.D
Acq On : 05 Oct 2023 13:27
Operator : JC/MD
Sample : VX1005MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1005MBL01

Quant Time: Oct 05 15:35:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

