

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100523\
 Data File : VX038104.D
 Acq On : 05 Oct 2023 16:31
 Operator : JC/MD
 Sample : 04742-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-TB-20230928

Quant Time: Oct 05 17:00:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	117432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204007	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92400	48.919	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.840%
35) Dibromofluoromethane	5.391	113	68242	46.423	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.840%
50) Toluene-d8	8.646	98	262628	46.754	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.500%
62) 4-Bromofluorobenzene	11.079	95	94492	44.372	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100523\
Data File : VX038104.D
Acq On : 05 Oct 2023 16:31
Operator : JC/MD
Sample : 04742-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-TB-20230928

Quant Time: Oct 05 17:00:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

