

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100523\
 Data File : VX038111.D
 Acq On : 05 Oct 2023 20:23
 Operator : JC/MD
 Sample : VX1005WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1005WBL02

Quant Time: Oct 06 06:34:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Sep 28 06:08:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	23303	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	132238	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	124051	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	58663	34.923	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	116.400%#
60) 4-Bromofluorobenzene	11.079	95	57386	28.377	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.600%
63) Toluene-d8	8.647	98	163919	30.610	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.033%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100523\
Data File : VX038111.D
Acq On : 05 Oct 2023 20:23
Operator : JC/MD
Sample : VX1005WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1005WBL02

Quant Time: Oct 06 06:34:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Sep 28 06:08:32 2023
Response via : Initial Calibration

