

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100620\
 Data File : VX018794.D
 Acq On : 06 Oct 2020 12:59
 Operator : JC/SP
 Sample : L4240-01DL 5X
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG3Q7DL

Quant Time: Oct 07 01:31:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 07 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	132598	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	125317	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	64679	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	35378	4.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.20%
7) Chloroethane-d5	1.70	69	32513	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.34	63	63045	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	4.54	46	127671	62.84	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.68%
24) Chloroform-d	5.14	84	92984	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	6.03	65	53515	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	6.04	84	166417	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
36) 1,2-Dichloropropane-d6	7.37	67	50075	5.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.20%
41) Toluene-d8	8.70	98	150871	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
45) trans-1,3-Dichloropropene-	8.99	79	22459	4.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.60%
48) 2-Hexanone-d5	9.43	63	100151	61.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.38%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	40229	5.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.80%
65) 1,2-Dichlorobenzene-d4	12.36	152	64237	5.38	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	1.64	94	473	0.112	ug/L	74
8) Chloroethane	1.75	64	5389	1.242	ug/L #	58
14) Carbon disulfide	2.55	76	2772	0.134	ug/L	97
15) Methyl Acetate	2.82	43	48481	17.783	ug/L #	48
16) Methylene chloride	2.83	84	3568	0.446	ug/L #	22
21) 2-Butanone	4.53	43	3839	2.377	ug/L #	53
22) cis-1,2-Dichloroethene	4.57	96	554	0.076	ug/L #	55
29) 1,1,1-Trichloroethane	5.46	97	9796	0.661	ug/L	99
30) Cyclohexane	5.54	56	65198	6.341	ug/L	98
33) Benzene	6.06	78	2097	0.080	ug/L	100
37) 1,2-Dichloropropane	7.37	63	6951	1.057	ug/L #	83
39) cis-1,3-Dichloropropene	8.37	75	1938	0.185	ug/L	95
42) Toluene	8.76	91	361540	12.590	ug/L	99
43) 1,3,5-Trimethylbenzene	11.49	105	87558	3.222	ug/L	95
44) 1,2,4-Trimethylbenzene	11.79	105	268259	9.725	ug/L	98
50) 2-Hexanone	9.43	43	12378	4.475	ug/L #	72
53) Chlorobenzene	10.12	112	8108	0.413	ug/L #	84

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100620\
 Data File : VX018794.D
 Acq On : 06 Oct 2020 12:59
 Operator : JC/SP
 Sample : L4240-01DL 5X
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG3Q7DL

Quant Time: Oct 07 01:31:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 07 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)


54) Ethylbenzene	10.23	91	36785	1.143	ug/L	98
55) m,p-xylene	10.34	106	50166	3.995	ug/L	97
56) o-xylene	10.68	106	19159	1.651	ug/L	93
57) Styrene	10.68	104	1305	0.066	ug/L #	1
58) Isopropylbenzene	11.00	105	14112	0.434	ug/L	96
63) 1,3-Dichlorobenzene	12.01	146	8341	0.531	ug/L	99
64) 1,4-Dichlorobenzene	12.08	146	36669	2.339	ug/L	95
66) 1,2-Dichlorobenzene	12.37	146	10733	0.731	ug/L	93
69) 1,2,4-trichlorobenzene	13.63	180	84962	8.457	ug/L	96
70) Naphthalene	13.82	128	14197	0.865	ug/L	97
71) 1,2,3-Trichlorobenzene	14.00	180	30315	3.354	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100620\
Data File : VX018794.D
Acq On : 06 Oct 2020 12:59
Operator : JC/SP
Sample : L4240-01DL 5X
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG3Q7DL

Quant Time: Oct 07 01:31:54 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 07 01:30:09 2020
Response via : Initial Calibration





#6

Bromomethane

Concen: 0.112 ug/L

RT: 1.64 min Scan# 90

Delta R.T. -0.01 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :


BG3Q7DL

Tot Ion: 94 Resp: 473

Ion Ratio Lower Upper




94 100

96 66.2 63.5 117.9



Abundance Ion 94.00 (93.70 to 94.70): VX0

Ion 96.00 (95.70 to 96.70): VX0



#8

Chloroethane

Concen: 1.242 ug/L

RT: 1.75 min Scan# 108

Delta R.T. 0.03 min

Lab File: VX018794.D


Acq: 06 Oct 2020 12:59

Tgt Ion: 64 Resp: 5389

Ion Ratio Lower Upper



64 100


66 7.6 21.5 39.9#



Abundance Ion 64.10 (63.80 to 64.80): VX0

Ion 66.05 (65.75 to 66.75): VX0





#14

Carbon disulfide

Concen: 0.134 ug/L

RT: 2.55 min Scan# 240

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :


BG3Q7DL

Tot Ion: 76 Resp: 2772

Ion Ratio Lower Upper




76 100

78 8.2 7.4 11.2



Abundance Ion 76.05 (75.75 to 76.75): VX0

Ion 78.00 (77.70 to 78.70): VX0



#15

Methyl Acetate

Concen: 17.783 ug/L

RT: 2.82 min Scan# 285

Delta R.T. 0.07 min

Lab File: VX018794.D


Acq: 06 Oct 2020 12:59

Tgt Ion: 43 Resp: 48481

Ion Ratio Lower Upper



43 100


74 0.2 21.9 32.9#



Abundance Ion 43.05 (42.75 to 43.75): VX0

Ion 74.05 (73.75 to 74.75): VX0









#16
Methvlene chloride
Concen: 0.446 ug/L
RT: 2.83 min Scan# 286
Delta R.T. 0.00 min
Lab File: VX018794.D
Acq: 06 Oct 2020 12:59




Instrument :
MSVOA_X
ClientSampleId :
BG3Q7DL


Tot Ion: 84 Resp: 3568
Ion Ratio Lower Upper
84 100
86 180.0 45.6 84.8#
49 74.3 85.8 159.3#



#21
2-Butanone
Concen: 2.377 ug/L
RT: 4.53 min Scan# 564
Delta R.T. -0.12 min
Lab File: VX018794.D
Acq: 06 Oct 2020 12:59

Tgt Ion: 43 Resp: 3839
Ion Ratio Lower Upper
43 100
72 1.2 12.4 37.2#





#22

cis-1,2-Dichloroethene

Concen: 0.076 ug/L

RT: 4.57 min Scan# 571

Delta R.T. 0.01 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleID :

BG3Q7DL


Tot Ion: 96 Resp: 554

Ion Ratio Lower Upper

96 100

61 72.4 86.2 160.0#




68 0.0 0.3 0.7#



Abundance Ion 96.00 (95.70 to 96.70): VX0

Ion 61.05 (60.75 to 61.75): VX0

Ion 68.15 (67.85 to 68.85): VX0



#29

1,1,1-Trichloroethane

Concen: 0.661 ug/L

RT: 5.46 min Scan# 717

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59


Tgt Ion: 97 Resp: 9796

Ion Ratio Lower Upper

97 100

99 64.1 51.3 76.9



61 42.5 33.3 49.9




Abundance Ion 96.95 (96.65 to 97.65): VX0

Ion 99.05 (98.75 to 99.75): VX0

Ion 61.05 (60.75 to 61.75): VX0









#30
Cyclohexane
Concen: 6.341 ug/L
RT: 5.54 min Scan# 731
Delta R.T. 0.00 min
Lab File: VX018794.D
Acq: 06 Oct 2020 12:59




Instrument :
MSVOA_X
ClientSampleId :
BG3Q7DL


Tot Ion: 56 Resp: 65198
Ion Ratio Lower Upper
56 100
69 31.8 26.0 39.0
84 90.7 73.8 110.6



#33
Benzene
Concen: 0.080 ug/L
RT: 6.06 min Scan# 815
Delta R.T. -0.05 min
Lab File: VX018794.D
Acq: 06 Oct 2020 12:59

Tgt Ion: 78 Resp: 2097









#37
 1,2-Dichloropropane
 Concen: 1.057 ug/L
 RT: 7.37 min Scan# 1030
 Delta R.T. -0.12 min
 Lab File: VX018794.D
 Acq: 06 Oct 2020 12:59




Instrument :
 MSVOA_X
 ClientSampleId :
 BG3Q7DL


Tot Ion: 63 Resp: 6951
 Ion Ratio Lower Upper
 63 100
 112 0.3 4.7 7.1#



#39
 cis-1,3-Dichloropropene
 Concen: 0.185 ug/L
 RT: 8.37 min Scan# 1195
 Delta R.T. -0.04 min
 Lab File: VX018794.D
 Acq: 06 Oct 2020 12:59

Tgt Ion: 75 Resp: 1938
 Ion Ratio Lower Upper
 75 100
 77 34.7 22.5 41.7





#42

Toluene

Concen: 12.590 ug/L

RT: 8.76 min Scan# 1259

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :


BG3Q7DL

Tot Ion: 91 Resp: 361540




Ion Ratio Lower Upper

91 100

92 58.4 40.3 74.9



Abundance Ion 91.15 (90.85 to 91.85): VX0



#43

1,3,5-Trimethylbenzene

Concen: 3.222 ug/L

RT: 11.49 min Scan# 1706

Delta R.T. 0.00 min

Lab File: VX018794.D


Acq: 06 Oct 2020 12:59

Tgt Ion:105 Resp: 87558



Ion Ratio Lower Upper


105 100

120 50.2 37.7 56.5



Abundance Ion 105.00 (104.70 to 105.70): V





#44

1,2,4-Trimethylbenzene

Concen: 9.725 ug/L

RT: 11.79 min Scan# 1756

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :


BG3Q7DL

Tot Ion:105 Resp: 268259

Ion Ratio Lower Upper




105 100

120 45.3 35.1 52.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#50

2-Hexanone

Concen: 4.475 ug/L

RT: 9.43 min Scan# 1368

Delta R.T. -0.04 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Tgt Ion: 43 Resp: 12378


Ion Ratio Lower Upper

43 100

58 28.7 41.0 61.4#

57 28.0 15.9 23.9#

100 1.2 10.1 15.1#






Abundance Ion 43.05 (42.75 to 43.75): VX0

Ion 58.05 (57.75 to 58.75): VX0

Ion 57.20 (56.90 to 57.90): VX0

Ion 100.15 (99.85 to 100.85): VX0





#53

Chlorobenzene

Concen: 0.413 ug/L

RT: 10.12 min Scan# 1482

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :

BG3Q7DL


Tot Ion:112 Resp: 8108

Ion Ratio Lower Upper

112 100

114 32.4 21.3 39.5




77 69.2 42.4 63.6#



Abundance Ion 112.10 (111.80 to 112.80): V

Ion 114.05 (113.75 to 114.75): V

Ion 77.10 (76.80 to 77.80): VX0



#54

Ethylbenzene

Concen: 1.143 ug/L

RT: 10.23 min Scan# 1500

Delta R.T. 0.00 min

Lab File: VX018794.D


Acq: 06 Oct 2020 12:59

Tgt Ion: 91 Resp: 36785

Ion Ratio Lower Upper



91 100


106 29.8 21.5 39.9



Abundance Ion 91.15 (90.85 to 91.85): VX0

Ion 106.15 (105.85 to 106.85): V





#55

m,p-xylene

Concen: 3.995 ug/L

RT: 10.34 min Scan# 1518

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :


BG3Q7DL

Tot Ion:106 Resp: 50166

Ion Ratio Lower Upper




106 100

91 208.9 143.4 266.2



Abundance Ion 106.15 (105.85 to 106.85): V

Ion 91.15 (90.85 to 91.85): VX0



#56

o-xylene

Concen: 1.651 ug/L

RT: 10.68 min Scan# 1574

Delta R.T. 0.00 min

Lab File: VX018794.D


Acq: 06 Oct 2020 12:59

Tgt Ion:106 Resp: 19159

Ion Ratio Lower Upper



106 100


91 207.9 153.4 284.8



Abundance Ion 106.15 (105.85 to 106.85): V

Ion 91.15 (90.85 to 91.85): VX0





#57

Stvrene

Concen: 0.066 ug/L

RT: 10.68 min Scan# 1574

Delta R.T. -0.01 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampled :


BG3Q7DL

Tot Ion:104 Resp: 1305

Ion Ratio Lower Upper




104 100

78 269.6 37.1 68.9#



Abundance Ion 104.10 (103.80 to 104.80): V

Ion 78.10 (77.80 to 78.80): VX0



#58

Isopropylbenzene

Concen: 0.434 ug/L

RT: 11.00 min Scan# 1626

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59


Tgt Ion:105 Resp: 14112

Ion Ratio Lower Upper

105 100

120 27.6 20.2 30.2



77 17.6 13.3 19.9




Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.10 (119.80 to 120.80): V

Ion 77.10 (76.80 to 77.80): VX0









#63
 1,3-Dichlorobenzene
 Concen: 0.531 ug/L
 RT: 12.01 min Scan# 1791
 Delta R.T. 0.00 min
 Lab File: VX018794.D
 Acq: 06 Oct 2020 12:59

Instrument :
 MSVOA_X
 ClientSampleId :
 BG3Q7DL

Tot Ion:146 Resp: 8341
 Ion Ratio Lower Upper
 146 100
 111 41.2 28.4 52.8
 148 66.4 45.9 85.3




Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 111.00 (110.70 to 111.70): V
 Ion 148.00 (147.70 to 148.70): V






#64
 1,4-Dichlorobenzene
 Concen: 2.339 ug/L
 RT: 12.08 min Scan# 1803
 Delta R.T. 0.00 min
 Lab File: VX018794.D
 Acq: 06 Oct 2020 12:59

Tgt Ion:146 Resp: 36669
 Ion Ratio Lower Upper
 146 100
 111 39.4 28.6 53.2
 148 57.8 44.2 82.0



Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 111.00 (110.70 to 111.70): V
 Ion 148.00 (147.70 to 148.70): V





#66

1,2-Dichlorobenzene

Concen: 0.731 ug/L

RT: 12.37 min Scan# 1851

Delta R.T. -0.01 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampled :

BG3Q7DL





Tot Ion:146 Resp: 10733

Ion Ratio Lower Upper

146 100

111 39.2 31.7 47.5

148 52.9 49.0 73.6



#69

1,2,4-trichlorobenzene

Concen: 8.457 ug/L

RT: 13.63 min Scan# 2057

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59




Tgt Ion:180 Resp: 84962


Ion Ratio Lower Upper

180 100

182 96.1 80.2 120.4

145 33.2 25.2 37.8





#70

Naphthalene

Concen: 0.865 ug/L

RT: 13.82 min Scan# 2088

Delta R.T. 0.00 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59

Instrument :

MSVOA_X

ClientSampleId :

BG3Q7DL


Tot Ion:128 Resp: 14197

Ion Ratio Lower Upper

128 100

129 12.4 9.0 13.4




127 13.9 10.5 15.7



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 127.00 (126.70 to 127.70): V



#71

1,2,3-Trichlorobenzene

Concen: 3.354 ug/L

RT: 14.00 min Scan# 2118

Delta R.T. -0.01 min

Lab File: VX018794.D

Acq: 06 Oct 2020 12:59


Tgt Ion:180 Resp: 30315

Ion Ratio Lower Upper

180 100

182 98.4 73.6 110.4

145 32.5 26.6 40.0



Abundance Ion 179.90 (179.60 to 180.60): V

Ion 181.90 (181.60 to 182.60): V

Ion 144.95 (144.65 to 145.65): V

