

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100620\
 Data File : VX018790.D
 Acq On : 06 Oct 2020 11:08
 Operator : JC/SP
 Sample : VX1006WBL01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK80

Quant Time: Oct 07 01:41:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 07 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	114608	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	114014	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	49070	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	29541	4.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.20%
7) Chloroethane-d5	1.70	69	26589	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.60%
11) 1,1-Dichloroethene-d2	2.34	63	53207	3.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.40%
20) 2-Butanone-d5	4.56	46	106518	60.66	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.32%
24) Chloroform-d	5.14	84	80053	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	6.03	65	46633	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	6.05	84	136145	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	7.37	67	42537	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	8.70	98	116039	4.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.00%
45) trans-1,3-Dichloropropene-	8.99	79	19292	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
48) 2-Hexanone-d5	9.43	63	82215	55.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.32%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	35021	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
65) 1,2-Dichlorobenzene-d4	12.36	152	47500	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100620\
Data File : VX018790.D
Acq On : 06 Oct 2020 11:08
Operator : JC/SP
Sample : VX1006WBL01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK80

Quant Time: Oct 07 01:41:03 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 07 01:30:09 2020
Response via : Initial Calibration

