

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX100620\
 Data File : VX018793.D
 Acq On : 06 Oct 2020 12:34
 Operator : JC/SP
 Sample : VIBLK61
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK61

Quant Time: Oct 07 01:56:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 07 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	134323	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	126433	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	60120	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	36256	4.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.40%
7) Chloroethane-d5	1.70	69	32279	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.34	63	65543	3.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.00%
20) 2-Butanone-d5	4.56	46	126765	61.59	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.18%
24) Chloroform-d	5.14	84	91915	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
26) 1,2-Dichloroethane-d4	6.03	65	53762	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	6.05	84	168581	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.40%
36) 1,2-Dichloropropane-d6	7.37	67	52511	5.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.60%
41) Toluene-d8	8.70	98	150037	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
45) trans-1,3-Dichloropropene-	8.99	79	22812	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
48) 2-Hexanone-d5	9.43	63	99154	60.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.08%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	40497	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
65) 1,2-Dichlorobenzene-d4	12.36	152	62536	5.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100620\
Data File : VX018793.D
Acq On : 06 Oct 2020 12:34
Operator : JC/SP
Sample : VIBLK61
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK61

Quant Time: Oct 07 01:56:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 07 01:30:09 2020
Response via : Initial Calibration

