

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
 Data File : VX024613.D
 Acq On : 06 Oct 2021 11:41
 Operator : JC/MD
 Sample : VX1006WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1006WBL01

Quant Time: Oct 06 15:21:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	146279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261964	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	249918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110266	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102890	49.907	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.820%
35) Dibromofluoromethane	5.397	113	85696	49.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	8.653	98	315089	49.193	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.380%
62) 4-Bromofluorobenzene	11.085	95	122098	48.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
Data File : VX024613.D
Acq On : 06 Oct 2021 11:41
Operator : JC/MD
Sample : VX1006WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1006WBL01

Quant Time: Oct 06 15:21:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

