

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
 Data File : VX024617.D
 Acq On : 06 Oct 2021 14:23
 Operator : JC/MD
 Sample : M4042-05
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-03-100121

Quant Time: Oct 06 15:23:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	235047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98962	46.959	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.920%
35) Dibromofluoromethane	5.391	113	80759	51.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.440%
50) Toluene-d8	8.653	98	331582	57.696	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	115.400%#
62) 4-Bromofluorobenzene	11.085	95	122893	54.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
Data File : VX024617.D
Acq On : 06 Oct 2021 14:23
Operator : JC/MD
Sample : M4042-05
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EO-03-100121

Quant Time: Oct 06 15:23:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

