

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
 Data File : VX024625.D
 Acq On : 06 Oct 2021 18:08
 Operator : JC/MD
 Sample : M4088-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-20211005-WATER

Quant Time: Oct 07 07:48:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96995	54.042	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.080%	
35) Dibromofluoromethane	5.397	113	75786	51.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.720%	
50) Toluene-d8	8.653	98	277602	51.114	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.220%	
62) 4-Bromofluorobenzene	11.085	95	104696	49.029	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.060%	
Target Compounds						
					Qvalue	
25) 2-Butanone	4.580	43	3387	2.740	ug/l	100
30) Chloroform	5.105	83	32880	12.297	ug/l	96
47) Bromodichloromethane	7.830	83	1682	0.779	ug/l	96
52) Toluene	8.720	92	1249	0.318	ug/l	90
68) m/p-Xylenes	10.305	106	1865	0.597	ug/l	99
69) o-Xylene	10.646	106	1018	0.329	ug/l	78
84) 1,2,4-Trimethylbenzene	11.756	105	5606	0.941	ug/l	96
95) Naphthalene	13.780	128	57123	8.251	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
Data File : VX024625.D
Acq On : 06 Oct 2021 18:08
Operator : JC/MD
Sample : M4088-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-20211005-WATER

Quant Time: Oct 07 07:48:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

