

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
 Data File : VX024627.D
 Acq On : 06 Oct 2021 18:55
 Operator : JC/MD
 Sample : M3960-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAMB-MW11-GW

Quant Time: Oct 07 07:48:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	146829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251465	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104932	50.707	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.420%
35) Dibromofluoromethane	5.397	113	86107	50.416	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.840%
50) Toluene-d8	8.653	98	316061	50.279	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.085	95	112808	45.642	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.280%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.947	59	9594	7.961	ug/l #	76
29) Tetrahydrofuran	5.032	42	1583	1.743	ug/l #	78
30) Chloroform	5.099	83	1504	0.488	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
Data File : VX024627.D
Acq On : 06 Oct 2021 18:55
Operator : JC/MD
Sample : M3960-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAMB-MW11-GW

Quant Time: Oct 07 07:48:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

