

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
 Data File : VX024628.D
 Acq On : 06 Oct 2021 19:19
 Operator : JC/MD
 Sample : M3960-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAMB-MW13-GWRE

Quant Time: Oct 07 07:49:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	147290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	206079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104624	50.400	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.800%		
35) Dibromofluoromethane	5.397	113	86001	50.917	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.840%		
50) Toluene-d8	8.653	98	268331	43.163	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 86.320%#		
62) 4-Bromofluorobenzene	11.085	95	100219	41.002	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 82.000%#		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	12404	4.191	ug/l	91
39) Methylcyclohexane	7.379	83	5975m	2.094	ug/l	
52) Toluene	8.720	92	1155	0.257	ug/l	83
67) Ethyl Benzene	10.195	91	28863	3.798	ug/l	95
68) m/p-Xylenes	10.305	106	35844	12.300	ug/l	97
69) o-Xylene	10.646	106	21274	7.382	ug/l	96
73) Isopropylbenzene	10.963	105	10735	1.557	ug/l	98
78) n-propylbenzene	11.305	91	10746	1.333	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	31375	5.325	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	48139	8.218	ug/l	97
95) Naphthalene	13.780	128	50389	7.401	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100621\
Data File : VX024628.D
Acq On : 06 Oct 2021 19:19
Operator : JC/MD
Sample : M3960-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAMB-MW13-GWRE

Quant Time: Oct 07 07:49:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

