

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031917.D
 Acq On : 06 Oct 2022 13:21
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 07 08:05:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	104412	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66410	47.902	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.800%
35) Dibromofluoromethane	5.385	113	56820	48.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.160%
50) Toluene-d8	8.647	98	205873	48.408	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.079	95	70689	47.017	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031917.D
Acq On : 06 Oct 2022 13:21
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 07 08:05:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

