

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031924.D
 Acq On : 06 Oct 2022 17:46
 Operator : JC/MD
 Sample : N4915-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 257155-3

Quant Time: Oct 07 08:07:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	64136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	106026	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69749	52.659	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.320%
35) Dibromofluoromethane	5.385	113	59622	49.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.653	98	211256	48.918	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	73070	47.860	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.720%
Target Compounds						
16) Acetone	2.386	43	1329	2.952	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031924.D
Acq On : 06 Oct 2022 17:46
Operator : JC/MD
Sample : N4915-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
257155-3

Quant Time: Oct 07 08:07:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

