

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031928.D
 Acq On : 06 Oct 2022 19:19
 Operator : JC/MD
 Sample : N5029-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-02-360-380-100522

Quant Time: Oct 07 08:08:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	61844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	100559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67408	52.778	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.560%			
35) Dibromofluoromethane	5.385	113	55259	48.546	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.100%			
50) Toluene-d8	8.647	98	200788	49.022	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.040%			
62) 4-Bromofluorobenzene	11.079	95	69741	48.163	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.320%			
Target Compounds						
16) Acetone	2.386	43	1923	4.430	ug/l #	74
52) Toluene	8.720	92	107398	37.164	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031928.D
Acq On : 06 Oct 2022 19:19
Operator : JC/MD
Sample : N5029-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-02-360-380-100522

Quant Time: Oct 07 08:08:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

