

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031929.D
 Acq On : 06 Oct 2022 19:42
 Operator : JC/MD
 Sample : N4882-15DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-127DL

Quant Time: Oct 07 08:08:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	58592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	98253	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64126	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67364	55.671	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.340%
35) Dibromofluoromethane	5.385	113	54886	49.350	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.653	98	198782	49.671	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	67684	47.840	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.680%
Target Compounds						
11) Tert butyl alcohol	2.977	59	20954	139.183	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	102101	34.092	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031929.D
Acq On : 06 Oct 2022 19:42
Operator : JC/MD
Sample : N4882-15DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-127DL

Quant Time: Oct 07 08:08:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

