

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031930.D
 Acq On : 06 Oct 2022 20:05
 Operator : JC/MD
 Sample : N4882-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-124

Quant Time: Oct 07 08:08:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	59717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	92821	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63148	51.203	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.400%			
35) Dibromofluoromethane	5.385	113	52210	49.692	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.380%			
50) Toluene-d8	8.647	98	190461	50.377	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.760%			
62) 4-Bromofluorobenzene	11.079	95	65457	48.973	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.940%			
Target Compounds						
16) Acetone	2.386	43	1098	2.620	ug/l #	83
20) Methylene Chloride	2.794	84	516	0.468	ug/l #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031930.D
Acq On : 06 Oct 2022 20:05
Operator : JC/MD
Sample : N4882-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-124

Quant Time: Oct 07 08:08:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

