

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031931.D
 Acq On : 06 Oct 2022 20:28
 Operator : JC/MD
 Sample : N4883-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB092822

Quant Time: Oct 07 08:08:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	57868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	92658	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63151	52.842	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.680%
35) Dibromofluoromethane	5.385	113	51592	49.190	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	8.653	98	189811	50.293	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	64240	48.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.300%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	817	2.011	ug/l #	77
20) Methylene Chloride	2.788	84	402	0.377	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031931.D
Acq On : 06 Oct 2022 20:28
Operator : JC/MD
Sample : N4883-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB092822

Quant Time: Oct 07 08:08:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

