

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031932.D
 Acq On : 06 Oct 2022 20:51
 Operator : JC/MD
 Sample : N4882-20 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-132

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/07/2022
 Supervised By : Mahesh Dadoda 10/07/2022

Quant Time: Oct 07 08:08:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	39891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	74199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	73827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62166	75.459	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 150.920%#			
35) Dibromofluoromethane	5.385	113	49704	59.179	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 118.360%			
50) Toluene-d8	8.647	98	139145	46.041	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.080%			
62) 4-Bromofluorobenzene	11.079	95	44835	41.963	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 83.920%			
Target Compounds						
11) Tert butyl alcohol	2.995	59	8833m	86.177	ug/l	
19) Methyl tert-butyl Ether	3.111	73	92552	45.391	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031932.D
Acq On : 06 Oct 2022 20:51
Operator : JC/MD
Sample : N4882-20 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-132

Quant Time: Oct 07 08:08:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/07/2022
Supervised By : Mahesh Dadoda 10/07/2022

