

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031936.D
 Acq On : 06 Oct 2022 22:24
 Operator : JC/MD
 Sample : N4882-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-129

Quant Time: Oct 07 08:09:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	54834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	90207	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61790	54.564	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.120%
35) Dibromofluoromethane	5.385	113	50949	49.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.800%
50) Toluene-d8	8.653	98	184848	50.309	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.079	95	62420	48.054	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.100%
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	15722	111.587	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	134963	48.153	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031936.D
 Acq On : 06 Oct 2022 22:24
 Operator : JC/MD
 Sample : N4882-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-129

Quant Time: Oct 07 08:09:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

